



**PROBEWELL**  
 4500 Rue Michelet  
 Quebec, QC, G1P0B5, Canada  
 Phone: 4186261126  
 Email: support@probewell.com  
 Website: probewell.com

**Work Order: 987654321**

**John Doe**  
 John.d@utility.com  
 Employee Id 123456



**Accuracy**



Test date: 02-13-2023 1:16:32 PM  
 Report date: 02-15-2023 9:47:45 AM  
 Last calibration: 02-13-2023  
 Serial number: 10000002  
 Tester model: MT-1/WT1

**Calibration Standard Used**

Manufacturer/Model	Serial number	Description	Calibrated
Radian	RD-21	208805	9-21-2022

**Watt Hour Collected Data**

Current	120.0V		240.0V		480.0V	
	Unity	60° Lag	Unity	60° Lag	Unity	60° Lag
1.5 A	-	-	-0.00350	-	-	-
3.0 A	-	-	0.00020	-	-	-
5.0 A	-	-	0.00340	-	-	-
15.0 A	-	-	-0.00020	-	-	-
30.0 A	-	-	0.00060	-	-	-
50.0 A	-	-	0.00360	-	-	-
Average	-	-	0.00	-	-	-
Max	-	-	0.0036	-	-	-
Min	-	-	-0.0035	-	-	-

**Overall**

AVERAGE	0.0007	N/A
MAXIMUM	0.0036	N/A
MINIMUM	-0.0035	N/A

**Accuracy specifications**

Typical	± 0.020
MAXIMUM	+ 0.050
MINIMUM	- 0.050