

# XT Accuracy Report Sample

Utility's Contact Info

Technician ID

Utility's logo



**Utility Name**  
 1234 Electric Avenue  
 Palm Beach, Florida, G1P0B5, United State  
 Phone: 418-626-1126  
 Email: support@utility.com  
 Website: utility.com

**John Doe**  
 john.doe@utility.com  
 Employee Id 1701

Test Type



Accuracy



Test date: 07-27-2023 9:33 AM  
 Report date: 07-27-2023 9:33 AM  
 Serial number: 20029270  
 Tester model: ST-3/XT3

Test Specifications

Type of Standard Being Used for the Accuracy Test

**Calibration Standard Used**

Manufacturer/Model Radian  
 Model RD21  
 Serial number 208805  
 Last calibration 9-9-2022

Results

	Meas	A	B	C	Avg	Unit
240V	Voltage	239.88	239.97	239.96	239.94	V
	Accuracy	0.05	0.01	0.01	0.02	%
1.0A	Current	1.00	1.00	1.00	1.00	A
	Accuracy	0.03	0.03	0.02	0.03	%
2.0A	Current	2.00	2.00	2.00	2.00	A
	Accuracy	0.02	0.02	0.02	0.02	%
5.0A	Current	5.01	5.01	5.01	5.01	A
	Accuracy	0.14	0.14	0.13	0.14	%
10.0A	Current	10.02	10.02	10.02	10.02	A
	Accuracy	0.24	0.25	0.20	0.23	%
<b>Accuracy</b>		<b>Avg</b>	<b>Min</b>	<b>Max</b>		<b>Unit</b>
	Current	0.10	0.25	0.02		%
	Voltage	0.03	0.05	0.01		%

Overall Average